

WPAN (Wireless Personal Area Network) Addendum



Manuals are available from Rice Lake Weighing Systems at www.ricelake.com/manuals

Warranty information is available at www.ricelake.com/warranties

This addendum applies to Healthweigh WPAN models.



NOTE: To verify the WPAN connection, Rice Lake Weighing Systems recommends downloading the ST BLE Toolbox from an application store and installing the application to a mobile device.

Follow the information below when attempting to connect to the WPAN system:

Connection Procedure:

1. The WPAN system and interface powers on, stabilizes and goes to zero. If the system remains at zero, it does not broadcast its presence and is hidden from the WPAN client scans.
2. When the system exits the zero state, due to a load being placed on the scale platform, the scale WPAN system starts advertising its presence and waits for a WPAN client to connect.
3. Once a connection is made, the scale WPAN system sends a notification to the connected WPAN client, which reads the data from the server. In order for the client to get the notification, the client must enable the notification on the desired characteristics, right after the connection is made. An additional notification is sent after 2.5 seconds from the moment the connection was made. This extra notification helps a Windows® WPAN client to get the notification on the correct time.
4. As long as the client remains connected, the server maintains the connection. The server and client remain connected even if the scale is returned to zero. If the scale system remains at zero for 15 seconds without a connection made, the scale WPAN system stops advertising its presence and the system becomes hidden from WPAN client scans.

Profile Description:

1. The broadcasted device name is "Healthweigh".
2. The pairing code is 123456, although it is not required for establishing a WPAN connection with the module.
3. The "WeightMeasurement" characteristic always holds the last weight that has been locked by the system.
4. Although the services and characteristics have 128bit UUIDs, which makes the services and characteristics "Custom", the service and its characteristics follow the official "WeightScaleService" scheme—found at: <https://www.bluetooth.com/specifications/specs/weight-scale-service-1-0/>.
For each GATT element, the corresponding official 16bit UUID is embedded in the 128bit UUID, in the third and fourth bytes.
5. A short description on the service structure and functionality. The official 16bit UUID is marked bold on the 128bit UUID, for each GATT element that's described here. Each GATT element UUID is the combination of the official UUID, embedded in the base UUID 0x1248 **XXXX** 84211000 8000 00805F9B34FA.



NOTE: Most of the features and fields in the services and characteristics are not used.

- A. The "WeightScaleService" UUID=0x1248 **181D** 84211000 8000 00805F9B34FA contains two characteristics:
 - i. "WeightScaleFeature" UUID=0x1248 **2A9E** 84211000 8000 00805F9B34FA
[org.bluetooth.characteristic.weight_scale_feature](https://www.bluetooth.com/specifications/specs/weight-scale-service-1-0/); this characteristic contains flags and switches, which allows the consumer of the service to know about features supported by this profile implementation; for example, the resolution of the weight measurement of the system is ReadOnly
 - ii. "Weight Measurement" UUID=0x1248 **2A9D** 84211000 8000 00805F9B34FA
[org.bluetooth.characteristic.weight_measurement](https://www.bluetooth.com/specifications/specs/weight-scale-service-1-0/); this characteristic holds the actual data that the service has to offer; it contains data that is relevant to a specific measurement; the weight, TimeStamp, UserID, BMI and Height are the data pieces that construct the entire characteristic; additionally, one byte is dedicated for flags at the start of the characteristic; for example, the first bit indicates the unit of measurement (Imperial/SI)



© Rice Lake Weighing Systems Content subject to change without notice.

230 W. Coleman St. • Rice Lake, WI 54868 • USA USA: 800-472-6703 • International: +1-715-234-9171